



neodisher[®] Mielclear



Rinse aid for the automated reprocessing of dental instruments



Liquid concentrate

Fields of application:

- Used in the rinsing step when reprocessing dental instruments, utensils and trays in dental practices, as well as in dental and orthodontic clinics

Performance spectrum:

- Achieves wetting of stainless steel surfaces, as well as plastics which are difficult to wet

Special properties:

- Neutralises carried-over alkali residues and alkaline waters

Application and dosage:

neodisher Mielclear is used in washer disinfectors for dental instruments. neodisher Mielclear is dosed automatically via integrated dosing devices in the final rinse water and is 0.2 - 0.4 ml/l.

Use suitable dosing devices.

Notes on application:

- For professional use only.
- Do not mix with other products.
- Rinse out dosing system including suction hose with water before changing product.
- The instructions given by the manufacturer of the washer disinfectant are to be observed.
- Reprocessing should comply with all ordinances pursuant to the regulations on medical devices and should be performed with appropriate validated processes.
- Please observe the reprocessing recommendations of the medical device manufacturers according to the requirements of the DIN EN ISO 17664.

Technical data:

pH-value	4.1 - 3.5 (0.2 - 0.4 ml/l, determined in deionised water, 20 °C)
Viscosity	< 50 mPa s (concentrate, 20 °C)
Density	approx. 1.1 g/cm ³ (20 °C)

Ingredients:

Ingredients according to Regulation (EC) No 648/2004 on detergents:
< 5 % non-ionic surfactants, phosphates

CE-mark:

neodisher Mielclear complies with European guidelines for medical devices.

If a serious incident occurs with the product, report it to the manufacturer and the relevant national authority.

Storage information:

The product is sensitive to frost. Always store at a temperature between 0 °C and 30 °C. Usable for 3 years when stored as recommended. For expiry date refer to the stamp mark on the label behind the hourglass symbol .

Hazard and precautionary statements:

For further safety information, see Safety Data Sheets. These are available e.g. at www.drweigert.com under the category "Service/Downloads".

Dispose only when container is empty and closed. For disposal of product residues, refer to the Safety Data Sheet.

MB 4119/3-1
Date of issue: 09/2020

With the above information, to our current knowledge we describe our product regarding safety necessities, but we do not involve any quality description or promise certain properties.